

Certificate of analyses

Product: EnzyPlus Lactose / D-Galactose
Art. No: EZS 784
Batch No: 400002
Expiry date: May 2005

Diffchamb certifies that this batch has been approved by the Quality Control Department and that it conforms to the following specifications:

Specific activity

β -Galactosidase	≥ 580 U/ml
Galactose DH	≥ 23 U/ml

pH value

Vial 1	6.4 – 6.8
Vial 3	8.4 – 8.8

Solubilisation time

Vial 1	≤ 3 min
--------	--------------

Performance test

target values	$\pm 5\%$
---------------	-----------

Date: 2004-06-16



Jörgen Plymouth
Quality Manager
Diffchamb AB

Diffchamb AB

F O Petersons Gata 32
SE-421 31 Västra Frölunda
www.diffchamb.com

Telephone +46-(0)31 742 33 50
Telefax +46-(0)31 58 33 70
E-mail diffchamb@diffchamb.se

Org.No. 556314-3303
VAT No. SE 556314330301

Bankgiro 896-5576
Postgiro 85 01 54-6



ISO9001
Certificate
No 11184